

## TECHNIQUES AND BIOMEDICAL APPLICATIONS OF FIELD-CYCLING MRI

David J. Lurie, Lionel M. Broche, Gareth R. Davies and Kerrin J. Pine

Biomedical Physics, School of Medicine & Dentistry, University of Aberdeen, Foresterhill,  
Aberdeen AB25 2ZD, United Kingdom

E-mail: d.lurie@abdn.ac.uk

www.ffc-mri.org

In fast field-cycling (FFC) NMR, the magnetic field is switched between levels during the pulse sequence; the “fast” in the name implies that the field is switched in a time short compared with the sample’s  $T_1$  relaxation time. FFC has been used for several decades for relaxometry measurements ( $T_1$  as a function of field strength) of small samples. In our laboratory we have combined FFC with MRI using a number of techniques [1].

An important application of FFC is imaging free radicals using Field-Cycled Proton-Electron Double-Resonance Imaging (FC-PEDRI). Irradiation of the free radical’s ESR causes a transfer of polarisation from electron spins to coupled nuclear spins (the Overhauser effect), resulting in a difference in image intensity. FFC allows the ESR irradiation to be applied at low field (so relatively low frequency, and low non-resonant absorption), while NMR signal detection and imaging is carried out at higher field, to preserve SNR. We have previously constructed two FC-PEDRI scanners [2,3].

Relaxometric MRI can also be achieved using FFC. The aim is to obtain spatially-resolved  $T_1$  versus magnetic field data, by collecting images at a range of evolution field strengths. We have recently demonstrated methods for implementing relaxometry on localised regions defined on a pilot image [4]. We have also shown that FFC relaxometry can detect the formation of cross-linked fibrin protein from fibrinogen *in vitro*, in a model of the blood clotting process [5]. This relies on  $^{14}\text{N}$ - $^1\text{H}$  cross-relaxation phenomena, also known as “quadrupole dips” in the  $T_1$ -dispersion plot [6].

We have recently built a single-magnet FFC-MRI system with signal detection at 0.5 T, and we are nearing completion of a whole-body human FFC-MRI scanner with detection at 0.2 T.

[1] Lurie D.J. *et al.*, Comptes Rendus Physique **11**, 136-148 (2010).

[2] Lurie D.J. *et al.*, Phys.Med.Biol. **43**, 1877-1886 (1998).

[3] Lurie D.J. *et al.*, Magn.Reson.Imaging **23**, 175-181 (2005).

[4] Pine K.J., Davies G.R. and Lurie D.J., Magn.Reson.Med. **63**, 1698–1702 (2010).

[5] Broche, L.M., Ismail, S.R., Booth, N.A. and Lurie, D.J., Magn.Reson.Med., In Press (2012).

[6] Kimmich R., Winter S., Nusser W. And Spohn K.H., J.Magn.Reson. **68**, 263-282 (1986).