

## Some simulation tools for fast field-cycling NMR and MRI instruments

7 June  
12:30pm

Pascal Fries<sup>1,2</sup>, E. Belorizky<sup>3</sup>, G. Ferrante<sup>4</sup>, G. Galuppini<sup>5</sup>, L. Magni<sup>5</sup>

<sup>1</sup>Univ. Grenoble Alpes, INAC, 38000 Grenoble, France, <sup>2</sup>CEA, INAC, MEM, 38054 Grenoble, France, <sup>3</sup>Univ. Grenoble Alpes, LIPHY, 38000 Grenoble, France, <sup>4</sup>Stelar s.r.l., 27035 Mede, PV, Italy, <sup>5</sup>ICDS Lab, University of Pavia, 27100 Pavia, Italy

Conventional NMR and MRI instruments operating at a fixed magnetic field have been enhanced for several decades by generations of skilled physicists and engineers who have worked out their optimal operating conditions. Fast field-cycling (FFC) instruments are not so advanced despite their continuous technical improvements since the beginning of their marketing or pioneer development [1,2]. Here, we present some simulation tools which aim at reproducing the operating conditions of FFC-NMR relaxometers and FFC-MRI imagers in order to determine the instrumental defects, and more generally the unwanted magnetic field perturbations, which can be corrected by post-processing and those which lead to a definitive degradation of the free induction decay and have to be suppressed. Particular attention is given to the time stability of the magnetic field.

### References:

- [1] G. Ferrante and S. Sykora, *Adv. Inorg. Chem.* 57, 405 (2005).
- [2] D. J. Lurie, S. Aime, S. Baroni, N. A. Booth, L. M. Broche, C.-H. Choi, G. R. Davies, S. Ismail, D. Ó hÓgáin, K. J. Pine, *C. R. Physique* 11, 136 (2010).